## Re-examination of Temperatures Referred to in the Thermal Conductivity and Thermal Diffusivity Data Measured by the Transient Hot-Wire Method

H. Watanabe
National Research Laboratory of Metrology
AIST
1-4, Umezono 1-chome, Tsukuba
Ibaraki 305-8563, Japan

Two temperatures referred to in the thermal conductivity and thermal diffusivity data, measured with the transient hot-wire technique, are re-examined. The temperature  $T_{\lambda}$  assigned for the thermal conductivity is derived as follows:

$$T_{\lambda} = T_{B} + \Delta \theta_{\lambda} = T_{B} + \{ (\chi + \phi)/\chi \} \Delta T_{m}, \quad \Delta T_{m} = (1/2) \{ \Delta T(t_{i}) + \Delta T(t_{f}) \}$$

$$\tag{1}$$

to be somewhat revised from the temperature  $T_{\rm m}$ = $T_{\rm B}$ + $\Delta T_{\rm m}$  used hitherto. Then, the temperature  $T_{\kappa}$  for the thermal diffusivity is newly introduced as follows:

$$T_{\kappa} = T_{\rm B} + \Delta \theta_{\kappa},$$

$$\Delta \theta_{\kappa} = (1/2) \{ (\chi + \phi)/(\chi - \phi) \} \{ (Q_0/4\pi\lambda_0)^{-1} \Delta T_{\rm m}^2 - (1/4) (Q_0/4\pi\lambda_0) (\ln t_{\rm f}/t_{\rm i})^2 \} + (Q_0/4\pi\lambda_0) \ln 4,$$
(2)

which differs from the initial temperature (i.e., the bath temperature  $T_{\rm B}$ ) conventionally used hitherto. Where,  $\chi$  and  $\phi$  are the temperature dependence of the thermal conductivity and thermal diffusivity, and subscripts ( $_{\rm i}$ ) and ( $_{\rm f}$ ) mean initial and final. The study is applied to the measurement of n-Heptane, n-Pentane, and iso-Pentane, and the validity is confirmed.